

**Human Interleukin-6 (IL-6) ELISA Kit
(Enzyme-Linked Immunosorbent Assay)**

Catalog Number: CRS-B001

Pack Size: 96 tests

IMPORTANT: Please carefully read this manual before performing your experiment.

For Research Use Only. Not For Use in Diagnostic or Therapeutic Procedures

INTENDED USE

The kit is developed for quantitative detection of IL-6 in human serum and cell culture supernates. It is intended for research use only (RUO).

BACKGROUND

Interleukin 6 (IL-6) is an interleukin that acts as both a pro-inflammatory cytokine and an anti-inflammatory myokine. It plays important roles in the acute phase reaction, inflammation, hematopoiesis, bone metabolism, and cancer progression.

PRINCIPLE OF THE ASSAY

This assay kit is used to measure the levels of human Interleukin-6 (IL-6) by employing a standard sandwich-ELISA format. The micro-plate in the kit has been pre-coated with Anti-IL-6 Antibody. Firstly, add the standard samples provided in kit and your samples to the plate, incubate and wash the wells. Then add the Biotin-Anti-IL-6 Antibody to the plate and form Antibody-antigen-biotinylated antibody complex, incubate and wash the wells. Next add Streptavidin-HRP to the plate, incubate and wash the wells. At last, load the substrate into the wells and monitor solution color from blue to yellow. The reaction is stopped by the addition of a stop solution and the intensity of the absorbance can be measured at 450nm and 630nm. The OD Value reflects the amount of IL-6 bound.

PRECAUTIONS

1. This kit is for research use only and is not for use in diagnostic or therapeutic applications.
2. The kit is suitable for cell supernatant, serum and plasma samples.
3. Do not use reagents past their expiration date.
4. Do not mix or substitute reagents with those from other kits or other lot number kits.
5. If samples generate values higher than the highest standard, dilute the samples with the appropriate calibrator diluent and repeat the assay. If cell supernatant samples need step dilution, except for the final dilution with diluent, other intermediate dilutions can be in cell culture medium.
6. Differences in test results can be caused by a variety of factors, including laboratory operator, pipette usage, plate washing technique, reaction time or temperature, and kit storage.

7. This kit is designed to remove or reduce some endogenous interference factors in biological samples, and not all possible influencing factors have been removed.

MATERIALS PROVIDED

Table1. Materials provided

Catalog	Components	Size (96 tests)	Format	Storage	
				Unopened	Opened
CRB001-C01	Pre-coated Anti-IL-6 Antibody Microplate	1 plate	Solid	2-8°C	2-8°C
CRB001-C02	Human IL-6 Standard	20 µg×2	Powder	2-8°C	-70°C
CRB001-C03	Biotin-Anti-IL-6 Antibody Con. Solution	100 µL	Liquid	2-8°C	2-8°C
CRB001-C04	Biotin-Antibody Dilution Buffer	8 mL	Liquid	2-8°C	2-8°C
CRB001-C05	Streptavidin-HRP Con. Solution	500 µL	Liquid	2-8°C, avoid light	2-8°C, avoid light
CRB001-C06	Streptavidin-HRP Dilution Buffer	15 mL	Liquid	2-8°C	2-8°C
CRB001-C07	20× Washing Buffer	50 mL	Liquid	2-8°C	2-8°C
CRB001-C08	1× Dilution Buffer	15 mL×2	Liquid	2-8°C	2-8°C
CRB001-C09	Substrate Solution	12 mL	Liquid	2-8°C, avoid light	2-8°C, avoid light
CRB001-C10	Stop Solution	6 mL	Liquid	2-8°C	2-8°C

KIT STORAGE AND EXPIRATION DATE

1. The unopened kit is stable for 12 months from the date of manufacture if stored at 2°C to 8°C.
2. The opened kit should be stored per Table 1. The shelf life is 30 days from the date of opening.

Note: a. Do not use reagents past their expiration date.

b. Find the expiration date on the outside packaging.

REAGENTS/EQUIPMENT NEEDED BUT NOT SUPPLIED

Single or multi-channel micropipettes and pipette tips: 10 µL, 300 µL, 1000 µL;

37°C Incubator;

Single or dual wavelength microplate reader with 450nm and 630nm filter;

Tubes: 1.5 mL, 10 mL;

Timer;

Reagent bottle;

Deionized or distilled water.

REAGENT PREPARATION

Bring all reagents and samples to room temperature (20°C-25°C) before use. If crystals have formed in buffer solution, place the sample in an 37°C incubator until the crystals have completely dissolved and bring the solution back to room temperature before use.

According to Table 2, prepare the provided lyophilized product into a storage solution with ultrapure water, dissolve at room temperature for 15 to 30 minutes, and mix by gently pipetting, avoiding vigorous shaking or vortexing. The reconstituted storage solution should be stored at -70°C. It is recommended that the number of freezing and thawing should not exceed 1 time, and the size of the aliquot should not be less than 5 µg.

Table 2. Preparation method

Catalog	Components	Size (96 T)	Storage solution concentration.	Reconstituted water Vol.
CRB001-C02	Human IL-6 Standard	20 µg	20 µg/mL	1000 µL

RECOMMENDED SAMPLE PREPARATION

1. Working Solution Preparation

1.1 Preparation of 1×Washing Buffer

Dilute 50 mL 20×Washing Buffer with deionized or ultrapure water to 1000 mL.

1.2 Preparation of Biotin-Anti-IL-6 Antibody Solution


Prepare Biotin-Anti-IL-6 Antibody Solution by diluting 30 μ L of Biotin-Anti-IL-6 Antibody Con. Solution into 6 mL Biotin-Antibody Dilution Buffer, mix gently well. The solution was freshly prepared just before use.

1.3 Preparation of IL-6 Streptavidin-HRP Solution

Prepare IL-6 Streptavidin-HRP Solution by diluting 240 μ L of IL-6 Streptavidin-HRP Con. Solution into 12 mL Streptavidin-HRP Dilution Buffer, mix gently well. The solution was freshly prepared just before use.

2. Preparation of Standard curve

The concentration of the reconstituted human IL-6 Standard (CRB001-C02) is 20 μ g/mL, prepare Cm by diluting 5 μ L of the reconstituted human IL-6 Standard into 995 μ L 1×Dilution Buffer, mix gently well. Label 7 tubes, one for each standard point: Std.-1, Std.-2, Std.-3, Std.-4, Std.-5, Std.-6, Std.-7. According to the following dilution scheme: 4 μ L IL-6 Cm + 996 μ L 1×Dilution Buffer. Shake gently to mix, labeled Std.-1 (Std.-1 = 400 pg/mL). Prepare 1:1 serial dilutions for the standard curve as follows: Pipette 500 μ L of 1×Dilution Buffer into each tube. Pipette 500 μ L of diluted standard (concentration of standard = 400 pg/mL) into the first tube, labeled Std.-2 (Std.-2 = 200 pg/mL), and mix. Pipette 500 μ L of this dilution into the second tube, labeled Std.-3, and mix thoroughly before the next transfer. 1×Dilution Buffer serves as blank.

Tubes/ Solution Code	Human IL-6 stock solution	Cm	Std.-1	Std.-2	Std.-3	Std.-4	Std.-5	Std.-6	Std.-7
Operating									
Solution Con.	20 µg/mL	100 ng/mL	400 pg/mL	200 pg/mL	100 pg/mL	50 pg/mL	25 pg/mL	12.5 pg/mL	6.25 pg/mL
Dilution Buffer Vol.		995 µL	996 µL	500 µL	500 µL	500 µL	500 µL	500 µL	500 µL

3. Add Samples and Biotin-Antibody Solution

Add 50 µL IL-6 Standard to each well, or add 50 µL samples to each well, finally add 50 µL Biotin-Anti-IL-6 Antibody Solution to each well. Seal the plate with microplate sealing film and incubate at room temperature (18-25 °C) for 1.0 h.

4. Washing

Remove the remaining solution by aspiration, add 300 µL of 1×Washing Buffer to each well, gently tap the plate for 1 min, remove any remaining 1×Washing Buffer: by aspirating or decanting, invert the plate and blot it against paper towels. Repeat the wash step above for five times.

5. Add IL-6 Streptavidin-HRP Solution

For all wells, add 100 µL IL-6 Streptavidin-HRP Solution. Seal the plate with microplate sealing film and incubate at room temperature (18-25 °C) for 30 min, avoid light.

6. Washing

Repeat step 4.

7. Substrate Reaction

Add 100 µL Substrate Solution to each well. Seal the plate with microplate sealing film and incubate at room temperature (18-25 °C) for 15 min, avoid light.

8. Termination

Add 50 µL Stop Solution to each well and tap the plate gently to allow thorough mixing.

Note: the color in the wells should change from blue to yellow.

9. Data Recording


Read the absorbance at 450nm and 630nm using UV/Vis microplate spectrophotometer.

Note: To reduce the background noise, subtract the readings at 630nm from the readings at 450nm.

CALCULATION OF RESULTS


1. Calculate the mean absorbance for each standard, control and sample and subtract average zero standard optical density (O.D.).
2. The standard curve is plotted with the standard concentration as x-axis and the calibrated absorbance value as y-axis. Four parameters logistic are used to draw the standard curve and calculate the sample concentration.
3. Normal range of Standard curve: $R^2 \geq 0.9900$.
4. Detection range: 6.25 pg/mL-400 pg/mL. If the OD value of the sample to be tested is higher than 400 pg/mL, the sample shall be diluted with dilution buffer and assay repeated. If the OD value of the sample to be tested is lower than 6.25 pg/mL, the sample should be reported.

QUICK GUID



TYPICAL DATA

The following data is for reference only. The sample concentration was calculated based on the results of the standard curve.



SENSITIVITY

The minimum detectable concentration of human IL-6 is 4.438 pg/mL. The minimum detectable concentration was determined by adding twice standard deviations to the mean optical density value of twenty zero standard replicates and calculating the corresponding concentration.

PRECISION

1. Intra-assay Precision

Three samples of known concentration were tested ten times on one plate to assess intra-assay precision.

2. Inter-assay Precision

Three samples of known concentration were tested on three different plates, ten replicates in each plate to assess inter-assay precision.

Sample	Intra-assay Precision			Inter-assay Precision		
	1	2	3	1	2	3
n	10	10	10	30	30	30
Mean (pg/mL)	772.00	624.00	457.20	839.47	588.80	438.27
SD	42.13	20.49	19.28	37.38	29.53	21.66
CV (%)	5.46	3.28	4.22	4.45	5.02	4.94

Note: The example data is for reference only.

RECOVERY

The serums were added with different concentrations of human IL-6, and the serum without human IL-6 was used as background to calculate the recovery rate. The Range of the recovery rate is 92.24%-114.49%, and the average recovery is 104.40%.

Sample Type	Average% Recovery	Range
Serum(n=5)	104.40	92.24-114.49%

LINEARITY

To assess the linearity of the assay, samples spiked with high concentrations of human IL-6 were serially diluted with calibrator diluent to produce samples with values within the dynamic range of the assay.

		Serum
1:2	Average Recovery (%)	108.23
	Range (%)	98.89-112.70
1:4	Average Recovery (%)	97.06
	Range (%)	91.96-98.45
1:8	Average Recovery (%)	106.97
	Range (%)	103.40-109.99

Note: The example data is for reference only.

SPECIFICITY

This assay recognizes natural and recombinant human IL-6. No cross-reactivity was observed when this kit was used to analyze the following recombinant cytokines.

Human	
IL-1 β	IL-10
IL-2	IL-15
IL-4	IL-21
IL-7	GM-CSF
IFN- γ	TNF-alpha

SAMPLE VALUES

80 human healthy sample were evaluated for the presence of human IL-6 in this assay. No medical histories were available for the donors used in this study.

Sample	n	Concentration (pg/mL)	Detection percentage (%)	Detection average Concentration (pg/mL)
Serum	80	n.d. – 17.64	81.25	12.13


Note: n.d. means that the concentration value cannot be detected which is lower than the sensitivity. The example data is for reference only.

CALIBRATION

This immunoassay is calibrated against a highly purified *E. coli*-expressed recombinant human IL-6 (89/548). Reference Reagent is calibrated by NIBSC/WHO in June 2020.

NIBSC/WHO (89/548) approximate value (U/mL) = 0.1 \times Human IL-6 value (pg/mL)

PLATE LAYOUT



Note: Blank is a Blank Dilution Buffer hole.

TROUBLESHOOTING GUIDE

Problem	Cause	Solution
Poor standard curve	* Inaccurate pipetting	* Check pipettes
Large CV	* Inaccurate pipetting * Air bubbles in wells	* Check pipettes * Remove bubbles in wells
High background	* Plate is insufficiently washed * Contaminated wash buffer	* Review the manual for proper wash. * Make fresh wash buffer
Very low readings across the plate	* Incorrect wavelengths * Insufficient development time	* Check filters/reader * Increase development time
Samples are reading too high, but standard curve looks fine	* Samples contain cytokine levels above assay range	* Dilute samples and run again
Drift	* Interrupted assay set-up * Reagents not at room temperature	* Assay set-up should be continuous - have all standards and samples prepared appropriately before commencement of the assay * Ensure that all reagents are at room temperature before pipetting into the wells unless otherwise instructed in the antibody inserts